

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100520\
 Data File : VN063880.D
 Acq On : 5 Oct 2020 18:56
 Operator : JC/MD
 Sample : PB132063ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132063ZHE#12

Manual Integrations
 APPROVED

MMDadoda
 10/6/2020 12:57:05 PM

Quant Time: Oct 06 03:31:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	418541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	438956	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	167545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	193169	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	
35) Dibromofluoromethane	7.55	113	138833	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	10.06	98	537262	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.38	95	204250	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
Target Compounds						
16) Acetone	3.78	43	10402	9.386	ug/l	91
18) Methyl Acetate	4.29	43	4723m	1.641	ug/l	
20) Methylene Chloride	4.51	84	9293	3.752	ug/l	92
43) Isopropyl Acetate	8.13	43	85075	12.243	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100520\
Data File : VN063880.D
Acq On : 5 Oct 2020 18:56
Operator : JC/MD
Sample : PB132063ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132063ZHE#12

Manual Integrations
APPROVED

MMDadoda
10/6/2020 12:57:05 PM

Quant Time: Oct 06 03:31:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

