


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100520\
 Data File : VN063858.D
 Acq On : 5 Oct 2020 9:34
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Title : SW846 8260
 Last Update : Thu Oct 01 01:46:38 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	13807	PASS
75	95	30	60	57.1	34997	PASS
95	95	100	100	100.0	61269	PASS
96	95	5	9	7.2	4430	PASS
173	174	0.00	2	0.6	289	PASS
174	95	50	100	78.9	48336	PASS
175	174	5	9	7.1	3441	PASS
176	174	95	101	96.1	46453	PASS
177	176	5	9	6.7	3096	PASS