

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100521\
 Data File : VN068803.D
 Acq On : 5 Oct 2021 14:47
 Operator : JC/MD
 Sample : M4054-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-033-IDWGW-20211001

Quant Time: Oct 05 15:53:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 29 08:56:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	60309	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	319860	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.746	117	276021	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	148908	32.03	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.77%
60) 4-Bromofluorobenzene	12.736	95	100197	23.24	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.47%
63) Toluene-d8	10.443	98	401629	32.19	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	107.30%
Target Compounds						
25) Chloroform	7.817	83	29837	3.97	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100521\
Data File : VN068803.D
Acq On : 5 Oct 2021 14:47
Operator : JC/MD
Sample : M4054-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-033-IDW/GW-20211001

Quant Time: Oct 05 15:53:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 29 08:56:32 2021
Response via : Initial Calibration

