

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074745.D
 Acq On : 05 Oct 2022 18:24
 Operator : JC\MD
 Sample : VN1005WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1005WBL01

Quant Time: Oct 06 07:39:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	305210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	578343	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	572874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	271828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	386993	53.513	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.020%
35) Dibromofluoromethane	8.177	113	306541	52.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.920%
50) Toluene-d8	10.571	98	1259900	50.249	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.853	95	328309	43.075	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074745.D
Acq On : 05 Oct 2022 18:24
Operator : JC\MD
Sample : VN1005WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1005WBL01

Quant Time: Oct 06 07:39:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

