

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074749.D
 Acq On : 05 Oct 2022 20:00
 Operator : JC\MD
 Sample : N4957-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-17

Quant Time: Oct 06 07:41:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	296664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	560202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	594967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	289808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	382644	54.436	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	108.880%	
35) Dibromofluoromethane	8.177	113	295410	52.192	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.380%	
50) Toluene-d8	10.571	98	1274632	52.483	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.960%	
62) 4-Bromofluorobenzene	12.847	95	344736	46.695	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.380%	
Target Compounds						
					Qvalue	
16) Acetone	4.465	43	135993	34.139	ug/l	95
20) Methylene Chloride	5.294	84	32944	3.116	ug/l #	82
43) Isopropyl Acetate	8.694	43	573552	25.257	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074749.D
Acq On : 05 Oct 2022 20:00
Operator : JC\MD
Sample : N4957-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-17

Quant Time: Oct 06 07:41:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

