

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074750.D
 Acq On : 05 Oct 2022 20:25
 Operator : JC\MD
 Sample : N4957-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-18

Quant Time: Oct 06 07:41:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	280036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	549966	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	564263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	281399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	361095	54.421	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.840%	
35) Dibromofluoromethane	8.177	113	291976	52.546	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.100%	
50) Toluene-d8	10.571	98	1201146	50.378	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.853	95	323966	44.698	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	136751	36.368	ug/l	92
43) Isopropyl Acetate	8.700	43	581667	26.091	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074750.D
Acq On : 05 Oct 2022 20:25
Operator : JC\MD
Sample : N4957-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-18

Quant Time: Oct 06 07:41:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

