

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074751.D
 Acq On : 05 Oct 2022 20:49
 Operator : JC\MD
 Sample : N4957-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-19

Quant Time: Oct 06 07:42:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	282155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	548449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	575696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	278263	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	368020	55.048	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.100%
35) Dibromofluoromethane	8.177	113	280749	50.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	10.571	98	1210675	50.918	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.840%
62) 4-Bromofluorobenzene	12.853	95	332533	46.007	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.020%
Target Compounds						
						Qvalue
16) Acetone	4.465	43	138043	36.436	ug/l	96
20) Methylene Chloride	5.301	84	34272	3.627	ug/l	95
43) Isopropyl Acetate	8.700	43	571878m	25.723	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074751.D
Acq On : 05 Oct 2022 20:49
Operator : JC\MD
Sample : N4957-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-19

Quant Time: Oct 06 07:42:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

