

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074757.D
 Acq On : 05 Oct 2022 23:14
 Operator : JC\MD
 Sample : N4957-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-25

Quant Time: Oct 06 07:44:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	276240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	513401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	550464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	273959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	359958	54.995	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.000%	
35) Dibromofluoromethane	8.177	113	276291	53.264	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.520%	
50) Toluene-d8	10.571	98	1195087	53.693	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	107.380%	
62) 4-Bromofluorobenzene	12.853	95	317186	46.880	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.471	43	128446	34.629	ug/l	92
18) Methyl Acetate	5.059	43	23377	1.702	ug/l	91
20) Methylene Chloride	5.289	84	26935	2.453	ug/l #	78
43) Isopropyl Acetate	8.700	43	552340	26.540	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074757.D
Acq On : 05 Oct 2022 23:14
Operator : JC\MD
Sample : N4957-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-25

Quant Time: Oct 06 07:44:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

