

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074759.D
 Acq On : 06 Oct 2022 00:02
 Operator : JC\MD
 Sample : N4957-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-27

Quant Time: Oct 06 07:45:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	262008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	497294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	528882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	258242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	352048	56.708	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.420%
35) Dibromofluoromethane	8.183	113	270797	53.896	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.800%
50) Toluene-d8	10.571	98	1112820	51.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.240%
62) 4-Bromofluorobenzene	12.853	95	301631	46.025	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.040%
Target Compounds						
						Qvalue
16) Acetone	4.465	43	126993	36.097	ug/l	94
43) Isopropyl Acetate	8.700	43	524676	26.027	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074759.D
Acq On : 06 Oct 2022 00:02
Operator : JC\MD
Sample : N4957-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-27

Quant Time: Oct 06 07:45:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

