

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074760.D
 Acq On : 06 Oct 2022 00:26
 Operator : JC\MD
 Sample : N4957-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-28

Quant Time: Oct 06 07:46:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	271934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	500666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	551356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	269558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	348464	54.082	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 108.160%		
35) Dibromofluoromethane	8.176	113	275646	54.492	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 108.980%		
50) Toluene-d8	10.570	98	1167092	53.769	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 107.540%		
62) 4-Bromofluorobenzene	12.853	95	317740	48.156	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 96.320%		
Target Compounds						Qvalue
16) Acetone	4.465	43	125768	34.444	ug/l	94
18) Methyl Acetate	5.065	43	30766	2.276	ug/l	95
43) Isopropyl Acetate	8.700	43	533206	26.272	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074760.D
Acq On : 06 Oct 2022 00:26
Operator : JC\MD
Sample : N4957-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-28

Quant Time: Oct 06 07:46:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

