

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074762.D
 Acq On : 06 Oct 2022 01:15
 Operator : JC\MD
 Sample : N4957-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-30

Quant Time: Oct 06 07:47:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	241503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	493307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	515503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	253092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	336514	58.808	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.620%	
35) Dibromofluoromethane	8.177	113	267832	53.737	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	107.480%	
50) Toluene-d8	10.570	98	1093208	51.117	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.240%	
62) 4-Bromofluorobenzene	12.853	95	296720	45.641	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	91.280%	
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	130217	40.156	ug/l	94
18) Methyl Acetate	5.059	43	19713	1.642	ug/l #	82
20) Methylene Chloride	5.300	84	27573	3.269	ug/l #	88
43) Isopropyl Acetate	8.700	43	543690	27.189	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074762.D
Acq On : 06 Oct 2022 01:15
Operator : JC\MD
Sample : N4957-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-30

Quant Time: Oct 06 07:47:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

