

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074763.D
 Acq On : 06 Oct 2022 01:39
 Operator : JC\MD
 Sample : N4957-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-31

Quant Time: Oct 06 07:47:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	261742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	478660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	533762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	257743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	339329	54.715	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.420%	
35) Dibromofluoromethane	8.177	113	253733	52.466	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.940%	
50) Toluene-d8	10.571	98	1118536	53.901	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.800%	
62) 4-Bromofluorobenzene	12.853	95	304839	48.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.640%	
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	144013	40.976	ug/l	97
18) Methyl Acetate	5.065	43	26096	2.006	ug/l	99
20) Methylene Chloride	5.300	84	26226	2.585	ug/l	87
43) Isopropyl Acetate	8.700	43	550016	28.347	ug/l #	91
52) Toluene	10.629	92	22312	1.466	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074763.D
Acq On : 06 Oct 2022 01:39
Operator : JC\MD
Sample : N4957-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-31

Quant Time: Oct 06 07:47:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

