

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074744.D
 Acq On : 05 Oct 2022 17:27
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 06 07:38:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	1.025	0.982	4.2	94	0.00
3 P	Chloromethane	1.525	1.335	12.5	94	0.00
4 C	Vinyl Chloride	1.888	1.723	8.7#	91	0.00
5 T	Bromomethane	1.331	1.278	4.0	92	0.01
6 T	Chloroethane	1.397	1.237	11.5	86	0.01
7 T	Trichlorofluoromethane	1.488	1.341	9.9	96	0.00
8 T	Diethyl Ether	0.776	0.696	10.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.912	0.855	6.3	97	0.00
10 T	Methyl Iodide	1.102	1.073	2.6	97	0.00
11 T	Tert butyl alcohol	0.397	0.329	17.1	88	0.00
12 CM	1,1-Dichloroethene	0.970	0.884	8.9#	99	0.00
13 T	Acrolein	0.112	0.094	16.1	91	0.00
14 T	Allyl chloride	1.994	1.899	4.8	94	0.00
15 T	Acrylonitrile	0.865	0.805	6.9	92	0.00
16 T	Acetone	0.671	0.616	8.2	94	0.00
17 T	Carbon Disulfide	2.652	2.441	8.0	95	0.00
18 T	Methyl Acetate	2.486	2.258	9.2	94	0.00
19 T	Methyl tert-butyl Ether	3.775	3.573	5.4	94	0.00
20 T	Methylene Chloride	1.263	1.126	10.8	98	0.00
21 T	trans-1,2-Dichloroethene	1.038	0.960	7.5	97	0.00
22 T	Diisopropyl ether	4.236	4.061	4.1	96	0.00
23 T	Vinyl Acetate	3.479	3.546	-1.9	96	0.00
24 P	1,1-Dichloroethane	2.184	2.068	5.3	95	0.00
25 T	2-Butanone	1.218	1.132	7.1	93	0.00
26 T	2,2-Dichloropropane	1.801	1.626	9.7	94	0.00
27 T	cis-1,2-Dichloroethene	1.236	1.192	3.6	97	0.00
28 T	Bromochloromethane	0.985	1.004	-1.9	94	0.00
29 T	Tetrahydrofuran	0.849	0.795	6.4	93	0.00
30 C	Chloroform	2.000	1.979	1.0#	98	0.00
31 T	Cyclohexane	2.048	1.939	5.3	96	0.00
32 T	1,1,1-Trichloroethane	1.705	1.678	1.6	96	0.00
33 S	1,2-Dichloroethane-d4	1.185	1.221	-3.0	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.505	0.517	-2.4	103	0.00
36 T	1,1-Dichloropropene	0.818	0.811	0.9	98	0.00
37 T	Ethyl Acetate	3.280	3.056	6.8	93	0.00
38 T	Carbon Tetrachloride	0.771	0.746	3.2	96	0.00
39 T	Methylcyclohexane	1.059	1.041	1.7	94	0.00
40 TM	Benzene	2.647	2.584	2.4	97	0.00
41 T	Methacrylonitrile	0.646	0.614	5.0	94	0.00
42 TM	1,2-Dichloroethane	0.831	0.815	1.9	94	0.00
43 T	Isopropyl Acetate	2.027	1.940	4.3	95	0.00
44 TM	Trichloroethene	0.551	0.533	3.3	96	0.00
45 C	1,2-Dichloropropane	0.717	0.701	2.2#	95	0.00
46 T	Dibromomethane	0.421	0.417	1.0	93	0.00
47 T	Bromodichloromethane	0.859	0.856	0.3	98	0.00
48 T	Methyl methacrylate	0.832	0.820	1.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
 Data File : VN074744.D
 Acq On : 05 Oct 2022 17:27
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 06 07:38:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.018	0.016	11.1	86	0.00
50 S	Toluene-d8	2.168	2.227	-2.7	100	0.00
51 T	4-Methyl-2-Pentanone	1.305	1.279	2.0	94	0.00
52 CM	Toluene	1.590	1.628	-2.4#	96	0.00
53 T	t-1,3-Dichloropropene	1.024	1.049	-2.4	96	0.00
54 T	cis-1,3-Dichloropropene	1.105	1.100	0.5	97	0.00
55 T	1,1,2-Trichloroethane	0.639	0.634	0.8	94	0.00
56 T	Ethyl methacrylate	1.165	1.180	-1.3	95	0.00
57 T	1,3-Dichloropropane	1.152	1.130	1.9	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.420	0.413	1.7	87	0.00
59 T	2-Hexanone	0.987	0.971	1.6	93	0.00
60 T	Dibromochloromethane	0.630	0.648	-2.9	96	0.00
61 T	1,2-Dibromoethane	0.650	0.647	0.5	96	0.00
62 S	4-Bromofluorobenzene	0.659	0.727	-10.3	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.406	0.370	8.9	95	0.00
65 PM	Chlorobenzene	1.532	1.497	2.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.559	0.529	5.4	96	0.00
67 C	Ethyl Benzene	2.855	2.845	0.4#	95	0.00
68 T	m/p-Xylenes	1.072	1.107	-3.3	98	0.00
69 T	o-Xylene	1.100	1.103	-0.3	96	0.00
70 T	Styrene	1.733	1.820	-5.0	97	0.00
71 P	Bromoform	0.427	0.435	-1.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.811	4.408	8.4	96	0.00
74 T	N-amyl acetate	2.909	2.462	15.4	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.915	1.606	16.1	97	0.00
76 T	1,2,3-Trichloropropane	1.492	1.418	5.0	96	0.00
77 T	Bromobenzene	1.013	0.888	12.3	96	0.00
78 T	n-propylbenzene	5.543	5.275	4.8	96	0.00
79 T	2-Chlorotoluene	3.361	3.047	9.3	97	0.00
80 T	1,3,5-Trimethylbenzene	3.837	3.677	4.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.766	0.666	13.1	91	0.00
82 T	4-Chlorotoluene	3.115	2.927	6.0	93	0.00
83 T	tert-Butylbenzene	3.316	3.149	5.0	95	0.00
84 T	1,2,4-Trimethylbenzene	3.880	3.698	4.7	96	0.00
85 T	sec-Butylbenzene	4.819	4.682	2.8	96	0.00
86 T	p-Isopropyltoluene	3.719	3.738	-0.5	96	0.00
87 T	1,3-Dichlorobenzene	1.908	1.721	9.8	93	0.00
88 T	1,4-Dichlorobenzene	1.802	1.753	2.7	97	0.00
89 T	n-Butylbenzene	3.278	3.304	-0.8	97	0.00
90 T	Hexachloroethane	0.831	0.757	8.9	96	0.00
91 T	1,2-Dichlorobenzene	1.838	1.795	2.3	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.399	0.335	16.0	95	0.00
93 T	1,2,4-Trichlorobenzene	0.804	0.769	4.4	95	0.00
94 T	Hexachlorobutadiene	0.376	0.325	13.6	97	0.00
95 T	Naphthalene	3.475	3.280	5.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100522\
Data File : VN074744.D
Acq On : 05 Oct 2022 17:27
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 06 07:38:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.875	0.749	14.4	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6