

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100618\
 Data File : VN051703.D
 Acq On : 5 Oct 2018 22:25
 Operator : MD\SY
 Sample : J5084-18ME
 Misc : 5.30g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 33 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 B-5(12-12.5)ME

Quant Time: Oct 09 08:00:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 05 02:22:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	608226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	896478	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	836034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	446183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	368470	45.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.72%	
35) Dibromofluoromethane	7.59	113	332895	48.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.06%	
50) Toluene-d8	10.09	98	1307501	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
62) 4-Bromofluorobenzene	12.40	95	438480	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
Target Compounds						
39) Methylcyclohexane	9.08	83	99720	8.478	ug/l #	78
67) Ethyl Benzene	11.51	91	30151	0.842	ug/l	94
73) Isopropylbenzene	12.25	105	49865	1.311	ug/l	99
78) n-propylbenzene	12.59	91	289620	6.981	ug/l	97
85) sec-Butylbenzene	13.17	105	148217	3.901	ug/l #	50
89) n-Butylbenzene	13.61	91	254010	9.539	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100618\
Data File : VN051703.D
Acq On : 5 Oct 2018 22:25
Operator : MD\SY
Sample : J5084-18ME
Misc : 5.30g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 33 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
B-5(12-12.5)ME

Quant Time: Oct 09 08:00:17 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 05 02:22:33 2018
Response via : Initial Calibration

