

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100618\
 Data File : VN051713.D
 Acq On : 6 Oct 2018 2:31
 Operator : MD\SY
 Sample : J5165-02 40X
 Misc : 6.24g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 43 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-092718-FD-S

Quant Time: Oct 06 04:08:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 05 02:22:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	690720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	994731	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	851657	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	312941	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	408753	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
35) Dibromofluoromethane	7.59	113	375781	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	10.09	98	1399355	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	12.40	95	387245	38.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.26%	
Target Compounds						
78) n-propylbenzene	12.60	91	11004	0.378	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	19634	0.872	ug/l	97
85) sec-Butylbenzene	13.17	105	9111	0.342	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100618\
Data File : VN051713.D
Acq On : 6 Oct 2018 2:31
Operator : MD\SY
Sample : J5165-02 40X
Misc : 6.24g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 43 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-092718-FD-S

Quant Time: Oct 06 04:08:18 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 05 02:22:33 2018
Response via : Initial Calibration

