

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100618\
 Data File : VN051717.D
 Acq On : 6 Oct 2018 4:10
 Operator : MD\SY
 Sample : J5165-06 40X
 Misc : 5.89g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 47 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-H-092718(7-10)

Manual Integrations
 APPROVED

MMDadoda
 10/8/2018 12:50:14 PM

Quant Time: Oct 06 06:06:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 05 02:22:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	684618	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	980208	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	859799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355039	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	402104	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
35) Dibromofluoromethane	7.59	113	372359	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	10.09	98	1384809	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	12.40	95	398262	39.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.60%	
Target Compounds						
7) Trichlorofluoromethane	2.87	101	64	Below Cal	#	19
41) Methacrylonitrile	7.18	41	658	Below Cal	#	25
73) Isopropylbenzene	12.25	105	21064	0.696	ug/l	95
78) n-propylbenzene	12.59	91	67349	2.040	ug/l	98
80) 1,3,5-Trimethylbenzene	12.73	105	37148	1.456	ug/l	96
83) tert-Butylbenzene	12.99	119	9353	0.409	ug/l	92
84) 1,2,4-Trimethylbenzene	13.04	105	541615	21.200	ug/l	100
85) sec-Butylbenzene	13.17	105	72255	2.390	ug/l	80
86) p-Isopropyltoluene	13.29	119	75812	2.829	ug/l	98
89) n-Butylbenzene	13.62	91	53431m	2.522	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100618\
Data File : VN051717.D
Acq On : 6 Oct 2018 4:10
Operator : MD\SY
Sample : J5165-06 40X
Misc : 5.89g/5.00mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 47 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-H-092718(7-10)

Manual Integrations
APPROVED

MMDadoda
10/8/2018 12:50:14 PM

Quant Time: Oct 06 06:06:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 05 02:22:33 2018
Response via : Initial Calibration

