


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100618\
Data File : VN051692.D
Acq On : 5 Oct 2018 17:23
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Title : SW846 8260
Last Update : Fri Oct 05 02:22:33 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	11218	PASS
75	95	30	60	56.4	28581	PASS
95	95	100	100	100.0	50672	PASS
96	95	5	9	6.3	3189	PASS
173	174	0.00	2	1.1	537	PASS
174	95	50	100	95.2	48253	PASS
175	174	5	9	7.7	3723	PASS
176	174	95	101	97.8	47210	PASS
177	176	5	9	6.9	3271	PASS