

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100620\
 Data File : VN063899.D
 Acq On : 6 Oct 2020 17:32
 Operator : JC/MD
 Sample : L4226-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-100120

Quant Time: Oct 06 18:05:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	240906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	435006	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	447461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	205960	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	7.55	113	149868	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
50) Toluene-d8	10.06	98	549747	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
62) 4-Bromofluorobenzene	12.38	95	215996	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
Target Compounds						
20) Methylene Chloride	4.50	84	408558	145.404	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100620\
Data File : VN063899.D
Acq On : 6 Oct 2020 17:32
Operator : JC/MD
Sample : L4226-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-100120

Quant Time: Oct 06 18:05:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

