

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100620\
 Data File : VN063901.D
 Acq On : 6 Oct 2020 18:20
 Operator : JC/MD
 Sample : L4221-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-100120

Quant Time: Oct 06 18:40:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	233680	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	419394	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	457602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	179296	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	199554	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
35) Dibromofluoromethane	7.55	113	147504	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.06	98	556797	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.38	95	217352	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
Target Compounds						
16) Acetone	3.78	43	14001	11.781	ug/l	92
20) Methylene Chloride	4.51	84	15420	5.752	ug/l	95
43) Isopropyl Acetate	8.13	43	72357	10.392	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100620\
Data File : VN063901.D
Acq On : 6 Oct 2020 18:20
Operator : JC/MD
Sample : L4221-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-100120

Quant Time: Oct 06 18:40:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

