

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100620\
 Data File : VN063921.D
 Acq On : 7 Oct 2020 3:29
 Operator : JC/MD
 Sample : L4242-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B17-100120-0-1

Manual Integrations
 APPROVED

MMDadoda
 10/7/2020 12:12:15 PM

Quant Time: Oct 07 05:59:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	224202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	422584	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	439684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	175897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	202977	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
35) Dibromofluoromethane	7.55	113	141433	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	10.06	98	551636	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	210883	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
Target Compounds						
16) Acetone	3.78	43	15939	13.979	ug/l	96
18) Methyl Acetate	4.29	43	3289m	1.111	ug/l	
20) Methylene Chloride	4.51	84	11878	4.637	ug/l	96
43) Isopropyl Acetate	8.13	43	86440	12.321	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100620\
Data File : VN063921.D
Acq On : 7 Oct 2020 3:29
Operator : JC/MD
Sample : L4242-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B17-100120-0-1

Manual Integrations
APPROVED

MMDadoda
10/7/2020 12:12:15 PM

Quant Time: Oct 07 05:59:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

