

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100620\
 Data File : VN063926.D
 Acq On : 7 Oct 2020 5:30
 Operator : JC/MD
 Sample : L4260-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B21-100220-0-1

Manual Integrations
 APPROVED

MMDadoda
 10/7/2020 12:12:23 PM

Quant Time: Oct 07 06:34:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	231624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	426515	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	449569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	181668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	204904	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
35) Dibromofluoromethane	7.55	113	146639	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.06	98	562198	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.38	95	215341	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
Target Compounds						
16) Acetone	3.78	43	13313	11.302	ug/l	92
18) Methyl Acetate	4.28	43	3544	1.159	ug/l #	77
20) Methylene Chloride	4.51	84	9802m	3.724	ug/l	
43) Isopropyl Acetate	8.13	43	88303	12.470	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100620\
Data File : VN063926.D
Acq On : 7 Oct 2020 5:30
Operator : JC/MD
Sample : L4260-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B21-100220-0-1

Manual Integrations
APPROVED

MMDadoda
10/7/2020 12:12:23 PM

Quant Time: Oct 07 06:34:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

