

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
 Data File : VN068820.D
 Acq On : 6 Oct 2021 15:52
 Operator : JC/MD
 Sample : VN1006WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1006WBL01

Quant Time: Oct 07 06:01:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	499703	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	854245	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	757404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	272678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	387674	52.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.58%	
35) Dibromofluoromethane	8.021	113	300097	53.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.96%	
50) Toluene-d8	10.443	98	1187159	54.25	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.50%	
62) 4-Bromofluorobenzene	12.733	95	343249	42.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
Data File : VN068820.D
Acq On : 6 Oct 2021 15:52
Operator : JC/MD
Sample : VN1006WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1006WBL01

Quant Time: Oct 07 06:01:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

