

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
 Data File : VN068826.D
 Acq On : 6 Oct 2021 18:40
 Operator : JC/MD
 Sample : M4072-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-207

Quant Time: Oct 07 06:02:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	503574	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	889974	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	857834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	298948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	394238	53.27	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.54%	
35) Dibromofluoromethane	8.027	113	306751	52.47	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.94%	
50) Toluene-d8	10.443	98	1251371	54.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.78%	
62) 4-Bromofluorobenzene	12.733	95	368007	43.40	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.80%	
Target Compounds						
						Qvalue
18) Methyl Acetate	4.867	43	9744	1.33	ug/l	# 76
20) Methylene Chloride	5.095	84	29711	3.27	ug/l	88
25) 2-Butanone	7.351	43	6185	1.83	ug/l	# 89
43) Isopropyl Acetate	8.566	43	59495	4.87	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
Data File : VN068826.D
Acq On : 6 Oct 2021 18:40
Operator : JC/MD
Sample : M4072-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-207

Quant Time: Oct 07 06:02:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

