

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
 Data File : VN068829.D
 Acq On : 6 Oct 2021 19:56
 Operator : JC/MD
 Sample : M4088-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-20211005-WATER

Quant Time: Oct 07 06:03:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	491034	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	878298	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	763626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	290799	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	388806	53.88	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.76%	
35) Dibromofluoromethane	8.024	113	301879	52.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.64%	
50) Toluene-d8	10.443	98	1206859	53.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.28%	
62) 4-Bromofluorobenzene	12.733	95	348600	41.86	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 83.72%	
Target Compounds						
						Qvalue
18) Methyl Acetate	4.867	43	9599	1.35	ug/l	# 84
25) 2-Butanone	7.345	43	6587	2.00	ug/l	91
30) Chloroform	7.817	83	90837	8.61	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100621\
Data File : VN068829.D
Acq On : 6 Oct 2021 19:56
Operator : JC/MD
Sample : M4088-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-20211005-WATER

Quant Time: Oct 07 06:03:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

