

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100621\  
 Data File : VN068821.D  
 Acq On : 6 Oct 2021 16:26  
 Operator : JC/MD  
 Sample : VN1006WBS01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1006WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 10/7/2021 5:34:27 PM

Quant Time: Oct 07 06:01:41 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N100621W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 06 14:41:09 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response           | Conc  | Units | Dev(Min) |
|------------------------------|----------------|------|--------------------|-------|-------|----------|
| -----                        |                |      |                    |       |       |          |
| Internal Standards           |                |      |                    |       |       |          |
| 1) Pentafluorobenzene        | 8.088          | 168  | 437125             | 50.00 | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene      | 8.971          | 114  | 750730             | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 11.746         | 117  | 709593             | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.677         | 152  | 304885             | 50.00 | ug/l  | 0.00     |
|                              |                |      |                    |       |       |          |
| System Monitoring Compounds  |                |      |                    |       |       |          |
| 33) 1,2-Dichloroethane-d4    | 8.440          | 65   | 327204             | 50.94 | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = 101.88% |       |       |          |
| 35) Dibromofluoromethane     | 8.024          | 113  | 253928             | 51.49 | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = 102.98% |       |       |          |
| 50) Toluene-d8               | 10.443         | 98   | 1009655            | 52.50 | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = 105.00% |       |       |          |
| 62) 4-Bromofluorobenzene     | 12.733         | 95   | 333436             | 46.26 | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = 92.52%  |       |       |          |
|                              |                |      |                    |       |       |          |
| Target Compounds             |                |      |                    |       |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.075          | 85   | 75430              | 19.02 | ug/l  | 96       |
| 3) Chloromethane             | 2.303          | 50   | 77034              | 16.79 | ug/l  | 98       |
| 4) Vinyl Chloride            | 2.445          | 62   | 142233             | 17.28 | ug/l  | 99       |
| 5) Bromomethane              | 2.858          | 94   | 144090             | 20.81 | ug/l  | 99       |
| 6) Chloroethane              | 3.019          | 64   | 128708             | 17.33 | ug/l  | 99       |
| 7) Trichlorofluoromethane    | 3.379          | 101  | 135728             | 19.13 | ug/l  | 100      |
| 8) Diethyl Ether             | 3.829          | 74   | 45645              | 19.88 | ug/l  | 86       |
| 9) 1,1,2-Trichlorotrifluo... | 4.215          | 101  | 78260              | 20.10 | ug/l  | 94       |
| 10) Methyl Iodide            | 4.425          | 142  | 82381              | 17.94 | ug/l  | # 89     |
| 11) Tert butyl alcohol       | 5.369          | 59   | 73184              | 95.84 | ug/l  | # 86     |
| 12) 1,1-Dichloroethene       | 4.189          | 96   | 66079              | 19.54 | ug/l  | 80       |
| 13) Acrolein                 | 4.044          | 56   | 35292              | 85.17 | ug/l  | 100      |
| 14) Allyl chloride           | 4.848          | 41   | 104677             | 18.61 | ug/l  | # 84     |
| 15) Acrylonitrile            | 5.556          | 53   | 190780             | 95.44 | ug/l  | 99       |
| 16) Acetone                  | 4.288          | 43   | 185342             | 90.96 | ug/l  | 92       |
| 17) Carbon Disulfide         | 4.537          | 76   | 124763             | 17.65 | ug/l  | 99       |
| 18) Methyl Acetate           | 4.859          | 43   | 123079             | 19.38 | ug/l  | 93       |
| 19) Methyl tert-butyl Ether  | 5.618          | 73   | 281897             | 19.57 | ug/l  | 96       |
| 20) Methylene Chloride       | 5.101          | 84   | 88834              | 19.51 | ug/l  | 88       |
| 21) trans-1,2-Dichloroethene | 5.602          | 96   | 72489              | 19.17 | ug/l  | 86       |
| 22) Diisopropyl ether        | 6.498          | 45   | 260824             | 19.10 | ug/l  | # 93     |
| 23) Vinyl Acetate            | 6.433          | 43   | 1008840            | 94.63 | ug/l  | 95       |
| 24) 1,1-Dichloroethane       | 6.391          | 63   | 152887             | 19.31 | ug/l  | 99       |
| 25) 2-Butanone               | 7.335          | 43   | 274086             | 93.47 | ug/l  | 92       |
| 26) 2,2-Dichloropropane      | 7.327          | 77   | 145760             | 19.23 | ug/l  | 95       |
| 27) cis-1,2-Dichloroethene   | 7.329          | 96   | 96785              | 18.89 | ug/l  | 86       |
| 28) Bromochloromethane       | 7.659          | 49   | 64761              | 15.05 | ug/l  | # 80     |
| 29) Tetrahydrofuran          | 7.683          | 42   | 168535             | 94.07 | ug/l  | 88       |
| 30) Chloroform               | 7.820          | 83   | 182691             | 19.46 | ug/l  | 98       |
| 31) Cyclohexane              | 8.099          | 56   | 125026             | 19.04 | ug/l  | 88       |
| 32) 1,1,1-Trichloroethane    | 8.016          | 97   | 160200             | 19.17 | ug/l  | 98       |
| 36) 1,1-Dichloropropene      | 8.222          | 75   | 113098             | 19.06 | ug/l  | 95       |
| 37) Ethyl Acetate            | 7.410          | 43   | 103639             | 18.29 | ug/l  | 99       |
| 38) Carbon Tetrachloride     | 8.209          | 117  | 135479             | 18.95 | ug/l  | 99       |
| 39) Methylcyclohexane        | 9.461          | 83   | 126036             | 18.97 | ug/l  | 91       |
| 40) Benzene                  | 8.464          | 78   | 360393             | 19.44 | ug/l  | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100621\  
 Data File : VN068821.D  
 Acq On : 6 Oct 2021 16:26  
 Operator : JC/MD  
 Sample : VN1006WBS01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1006WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 10/7/2021 5:34:27 PM

Quant Time: Oct 07 06:01:41 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N100621W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 06 14:41:09 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 7.640  | 41   | 53411    | 18.84  | ug/l # | 87       |
| 42) 1,2-Dichloroethane        | 8.531  | 62   | 141991   | 19.28  | ug/l   | 98       |
| 43) Isopropyl Acetate         | 8.560  | 43   | 202076   | 19.62  | ug/l   | 97       |
| 44) Trichloroethene           | 9.217  | 130  | 93674    | 19.51  | ug/l   | 86       |
| 45) 1,2-Dichloropropane       | 9.491  | 63   | 95565    | 19.88  | ug/l   | 99       |
| 46) Dibromomethane            | 9.577  | 93   | 70852    | 18.96  | ug/l   | 94       |
| 47) Bromodichloromethane      | 9.764  | 83   | 148773   | 19.82  | ug/l   | 98       |
| 48) Methyl methacrylate       | 9.561  | 41   | 84162    | 19.13  | ug/l # | 80       |
| 49) 1,4-Dioxane               | 9.569  | 88   | 37210    | 388.62 | ug/l # | 84       |
| 51) 4-Methyl-2-Pentanone      | 10.330 | 43   | 617584   | 94.18  | ug/l   | 97       |
| 52) Toluene                   | 10.507 | 92   | 239106   | 19.17  | ug/l   | 95       |
| 53) t-1,3-Dichloropropene     | 10.719 | 75   | 150952   | 19.67  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 10.191 | 75   | 155203   | 19.66  | ug/l   | 94       |
| 55) 1,1,2-Trichloroethane     | 10.899 | 97   | 103523   | 19.90  | ug/l   | 94       |
| 56) Ethyl methacrylate        | 10.762 | 69   | 138946   | 19.02  | ug/l   | 91       |
| 57) 1,3-Dichloropropane       | 11.046 | 76   | 164295   | 19.69  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 10.043 | 63   | 156981   | 101.70 | ug/l   | 92       |
| 59) 2-Hexanone                | 11.087 | 43   | 433872   | 91.74  | ug/l   | 97       |
| 60) Dibromochloromethane      | 11.240 | 129  | 112161   | 19.23  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.347 | 107  | 99778    | 18.86  | ug/l   | 98       |
| 64) Tetrachloroethene         | 10.979 | 164  | 82923    | 19.54  | ug/l   | 95       |
| 65) Chlorobenzene             | 11.771 | 112  | 270482   | 19.63  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 11.843 | 131  | 104638   | 19.38  | ug/l   | 98       |
| 67) Ethyl Benzene             | 11.846 | 91   | 475568   | 19.26  | ug/l   | 96       |
| 68) m/p-Xylenes               | 11.956 | 106  | 367806   | 39.18  | ug/l   | 85       |
| 69) o-Xylene                  | 12.283 | 106  | 183253   | 19.60  | ug/l   | 86       |
| 70) Styrene                   | 12.296 | 104  | 296530   | 19.05  | ug/l   | 93       |
| 71) Bromoform                 | 12.460 | 173  | 74722    | 18.84  | ug/l # | 97       |
| 73) Isopropylbenzene          | 12.581 | 105  | 485234   | 20.15  | ug/l   | 99       |
| 74) N-amyl acetate            | 12.390 | 43   | 152926   | 18.76  | ug/l   | 94       |
| 75) 1,1,2,2-Tetrachloroethane | 12.830 | 83   | 146532   | 20.19  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 12.881 | 75   | 119188m  | 19.83  | ug/l   |          |
| 77) Bromobenzene              | 12.862 | 156  | 109127   | 19.90  | ug/l   | 73       |
| 78) n-propylbenzene           | 12.924 | 91   | 560381   | 20.45  | ug/l   | 97       |
| 79) 2-Chlorotoluene           | 13.010 | 91   | 328935   | 20.16  | ug/l   | 90       |
| 80) 1,3,5-Trimethylbenzene    | 13.061 | 105  | 403440   | 20.25  | ug/l   | 95       |
| 81) trans-1,4-Dichloro-2-b... | 12.629 | 75   | 40073    | 19.32  | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 13.106 | 91   | 325586   | 20.11  | ug/l   | 93       |
| 83) tert-Butylbenzene         | 13.326 | 119  | 350463   | 19.91  | ug/l   | 91       |
| 84) 1,2,4-Trimethylbenzene    | 13.369 | 105  | 403028   | 20.61  | ug/l   | 96       |
| 85) sec-Butylbenzene          | 13.503 | 105  | 508024   | 20.28  | ug/l   | 96       |
| 86) p-Isopropyltoluene        | 13.618 | 119  | 412740   | 20.35  | ug/l   | 94       |
| 87) 1,3-Dichlorobenzene       | 13.618 | 146  | 202981   | 19.72  | ug/l   | 95       |
| 88) 1,4-Dichlorobenzene       | 13.696 | 146  | 197655   | 19.46  | ug/l   | 96       |
| 89) n-Butylbenzene            | 13.943 | 91   | 334738   | 19.56  | ug/l   | 96       |
| 90) Hexachloroethane          | 14.211 | 117  | 79901    | 20.38  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 13.991 | 146  | 198136   | 20.00  | ug/l   | 95       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.603 | 75   | 25016    | 20.18  | ug/l   | 74       |
| 93) 1,2,4-Trichlorobenzene    | 15.265 | 180  | 82083    | 18.10  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 15.372 | 225  | 49568    | 20.46  | ug/l   | 98       |
| 95) Naphthalene               | 15.509 | 128  | 200732   | 16.59  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.702 | 180  | 78390    | 18.80  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100621\  
Data File : VN068821.D  
Acq On : 6 Oct 2021 16:26  
Operator : JC/MD  
Sample : VN1006WBS01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM

Quant Time: Oct 07 06:01:41 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N100621W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 06 14:41:09 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#1  
Pentafluorobenzene  
Concen: 50.00 ug/l  
RT: 8.088 min Scan# 2274  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 168 Resp: 437125  
Ion Ratio Lower Upper  
168 100  
99 60.7 62.9 94.3#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#2  
Dichlorodifluoromethane  
Concen: 19.02 ug/l  
RT: 2.075 min Scan# 32  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 85 Resp: 75430  
Ion Ratio Lower Upper  
85 100  
87 30.0 16.1 48.2









#3  
Chloromethane  
Concen: 16.79 ug/l  
RT: 2.303 min Scan# 117  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 50 Resp: 77034  
Ion Ratio Lower Upper  
50 100  
52 33.2 25.6 38.4



#4  
Vinyl Chloride  
Concen: 17.28 ug/l  
RT: 2.445 min Scan# 170  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 62 Resp: 142233  
Ion Ratio Lower Upper  
62 100  
64 29.7 24.4 36.6




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM



#5  
Bromomethane  
Concen: 20.81 ug/l  
RT: 2.858 min Scan# 324  
Delta R.T. 0.010 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 94 Resp: 144090  
Ion Ratio Lower Upper  
94 100  
96 91.0 73.7 110.5




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#6  
Chloroethane  
Concen: 17.33 ug/l  
RT: 3.019 min Scan# 384  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 64 Resp: 128708  
Ion Ratio Lower Upper  
64 100  
66 31.5 24.7 37.1









#7  
Trichlorofluoromethane  
Concen: 19.13 ug/l  
RT: 3.379 min Scan# 518  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:101 Resp: 135728  
Ion Ratio Lower Upper  
101 100  
103 64.2 51.2 76.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#8  
Diethyl Ether  
Concen: 19.88 ug/l  
RT: 3.829 min Scan# 686  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 74 Resp: 45645  
Ion Ratio Lower Upper  
74 100  
45 86.3 50.2 150.7









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.10 ug/l  
RT: 4.215 min Scan# 830  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion:101 Resp: 78260  
Ion Ratio Lower Upper  
101 100  
85 49.1 39.6 59.4  
151 76.1 54.1 81.1



#10  
Methyl Iodide  
Concen: 17.94 ug/l  
RT: 4.425 min Scan# 908  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion:142 Resp: 82381  
Ion Ratio Lower Upper  
142 100  
127 47.8 43.1 64.7  
141 6.8 11.4 17.0#



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM







#11  
Tert butyl alcohol  
Concen: 95.84 ug/l  
RT: 5.369 min Scan# 1260  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 59 Resp: 73184  
Ion Ratio Lower Upper  
59 100  
57 5.4 8.7 13.1#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#12  
1,1-Dichloroethene  
Concen: 19.54 ug/l  
RT: 4.189 min Scan# 820  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 96 Resp: 66079  
Ion Ratio Lower Upper  
96 100  
61 153.0 152.6 229.0  
98 61.9 50.9 76.3









#13  
Acrolein  
Concen: 85.17 ug/l  
RT: 4.044 min Scan# 766  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 56 Resp: 35292  
Ion Ratio Lower Upper  
56 100  
55 71.6 57.4 86.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#14  
Allyl chloride  
Concen: 18.61 ug/l  
RT: 4.848 min Scan# 1066  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 41 Resp: 104677  
Ion Ratio Lower Upper  
41 100  
39 89.3 60.9 91.3  
76 39.6 24.3 36.5#









#15  
Acrylonitrile  
Concen: 95.44 ug/l  
RT: 5.556 min Scan# 1330  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 53 Resp: 190780  
Ion Ratio Lower Upper  
53 100  
52 83.6 66.2 99.2  
51 36.6 30.2 45.4

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#16  
Acetone  
Concen: 90.96 ug/l  
RT: 4.288 min Scan# 857  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 43 Resp: 185342  
Ion Ratio Lower Upper  
43 100  
58 30.5 21.2 31.8







#17  
Carbon Disulfide  
Concen: 17.65 ug/l  
RT: 4.537 min Scan# 950  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 76 Resp: 124763  
Ion Ratio Lower Upper  
76 100  
78 9.9 7.7 11.5




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#18  
Methyl Acetate  
Concen: 19.38 ug/l  
RT: 4.859 min Scan# 1070  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 43 Resp: 123079  
Ion Ratio Lower Upper  
43 100  
74 25.0 17.4 26.0





#19





Methyl tert-butyl Ether  
Concen: 19.57 ug/l  
RT: 5.618 min Scan# 1353  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 73 Resp: 281897  
Ion Ratio Lower Upper  
73 100  
57 19.2 17.0 25.6

Instrument :  
MSVOA\_N  
ClientSampled :  
VN1006WBS01

Manual Integrations  
APPROVED




MMDadoda  
10/7/2021 5:34:27 PM




#20

Methylene Chloride  
Concen: 19.51 ug/l  
RT: 5.101 min Scan# 1160  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 84 Resp: 88834  
Ion Ratio Lower Upper  
84 100  
49 119.1 110.9 166.3  
51 35.8 32.6 49.0  
86 58.5 51.6 77.4









#21  
trans-1,2-Dichloroethene  
Concen: 19.17 ug/l  
RT: 5.602 min Scan# 1347  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 96 Resp: 72489  
Ion Ratio Lower Upper  
96 100  
61 132.2 123.5 185.3  
98 64.1 48.6 72.8

Instrument :  
MSVOA\_N  
ClientSampled :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#22  
Diisopropyl ether  
Concen: 19.10 ug/l  
RT: 6.498 min Scan# 1681  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 45 Resp: 260824  
Ion Ratio Lower Upper  
45 100  
43 65.3 49.4 74.2  
87 32.3 20.9 31.3#  
59 13.8 9.0 13.4#









#23  
 Vinyl Acetate  
 Concen: 94.63 ug/l  
 RT: 6.433 min Scan# 1657  
 Delta R.T. -0.003 min  
 Lab File: VN068821.D  
 Acq: 6 Oct 2021 16:26

Tgt Ion: 43 Resp: 1008840  
 Ion Ratio Lower Upper  
 43 100  
 86 10.9 7.3 10.9

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1006WBS01




Manual Integrations  
 APPROVED


MMDadoda  
 10/7/2021 5:34:27 PM



#24  
 1,1-Dichloroethane  
 Concen: 19.31 ug/l  
 RT: 6.391 min Scan# 1641  
 Delta R.T. 0.003 min  
 Lab File: VN068821.D  
 Acq: 6 Oct 2021 16:26

Tgt Ion: 63 Resp: 152887  
 Ion Ratio Lower Upper  
 63 100  
 98 6.5 3.0 9.0  
 100 4.2 1.9 5.7









#25  
2-Butanone  
Concen: 93.47 ug/l  
RT: 7.335 min Scan# 1993  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 43 Resp: 274086  
Ion Ratio Lower Upper  
43 100  
72 28.7 19.7 29.5

Instrument :  
MSVOA\_N  
ClientSampled :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#26  
2,2-Dichloropropane  
Concen: 19.23 ug/l  
RT: 7.327 min Scan# 1990  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 77 Resp: 145760  
Ion Ratio Lower Upper  
77 100  
97 21.9 9.8 29.3









#27  
cis-1,2-Dichloroethene  
Concen: 18.89 ug/l  
RT: 7.329 min Scan# 1991  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 96 Resp: 96785  
Ion Ratio Lower Upper  
96 100  
61 143.5 0.0 338.8  
98 65.4 0.0 129.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#28  
Bromochloromethane  
Concen: 15.05 ug/l  
RT: 7.659 min Scan# 2114  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 49 Resp: 64761  
Ion Ratio Lower Upper  
49 100  
129 2.5 0.0 3.8  
130 77.8 49.7 74.5#









#29  
Tetrahydrofuran  
Concen: 94.07 ug/l  
RT: 7.683 min Scan# 2123  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 42 Resp: 168535  
Ion Ratio Lower Upper  
42 100  
72 49.1 32.8 49.2  
71 45.1 31.2 46.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#30  
Chloroform  
Concen: 19.46 ug/l  
RT: 7.820 min Scan# 2174  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 83 Resp: 182691  
Ion Ratio Lower Upper  
83 100  
85 66.0 51.5 77.3









#31  
Cyclohexane  
Concen: 19.04 ug/l  
RT: 8.099 min Scan# 2278  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 56 Resp: 125026  
Ion Ratio Lower Upper  
56 100  
69 38.2 27.3 40.9  
84 93.2 64.9 97.3



#32  
1,1,1-Trichloroethane  
Concen: 19.17 ug/l  
RT: 8.016 min Scan# 2247  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 97 Resp: 160200  
Ion Ratio Lower Upper  
97 100  
99 64.0 50.2 75.2  
61 46.8 38.2 57.2




Instrument :  
MSVOA\_N  
ClientSampled :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM



#33  
1,2-Dichloroethane-d4  
Concen: 50.94 ug/l  
RT: 8.440 min Scan# 2405  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 65 Resp: 327204  
Ion Ratio Lower Upper  
65 100  
67 50.8 0.0 98.6




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#34  
1,4-Difluorobenzene  
Concen: 50.00 ug/l  
RT: 8.971 min Scan# 2603  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 114 Resp: 750730  
Ion Ratio Lower Upper  
114 100  
63 19.8 0.0 49.0  
88 14.8 0.0 36.6









#35  
Dibromofluoromethane  
Concen: 51.49 ug/l  
RT: 8.024 min Scan# 2250  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 113 Resp: 253928  
Ion Ratio Lower Upper  
113 100  
111 102.6 83.8 125.8  
192 16.7 13.4 20.0



#36  
1,1-Dichloropropene  
Concen: 19.06 ug/l  
RT: 8.222 min Scan# 2324  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 75 Resp: 113098  
Ion Ratio Lower Upper  
75 100  
110 37.2 15.7 47.0  
77 30.9 24.6 37.0




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED




MMDadoda  
10/7/2021 5:34:27 PM




#37  
Ethyl Acetate  
Concen: 18.29 ug/l  
RT: 7.410 min Scan# 2021  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 43 Resp: 103639  
Ion Ratio Lower Upper  
43 100  
61 13.5 10.6 16.0  
70 9.7 8.1 12.1



#38  
Carbon Tetrachloride  
Concen: 18.95 ug/l  
RT: 8.209 min Scan# 2319  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 117 Resp: 135479  
Ion Ratio Lower Upper  
117 100  
119 94.0 75.4 113.2  
121 31.2 23.5 35.3



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM







#39  
Methylcyclohexane  
Concen: 18.97 ug/l  
RT: 9.461 min Scan# 2786  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 83 Resp: 126036  
Ion Ratio Lower Upper  
83 100  
55 72.3 66.6 99.8  
98 45.3 35.3 52.9

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#40  
Benzene  
Concen: 19.44 ug/l  
RT: 8.464 min Scan# 2414  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 78 Resp: 360393  
Ion Ratio Lower Upper  
78 100  
77 24.5 18.9 28.3






#41  
Methacrylonitrile  
Concen: 18.84 ug/l  
RT: 7.640 min Scan# 2107  
Delta R.T. 0.005 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 41 Resp: 53411


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 39  | 63.6  | 47.6  | 71.4  |
| 67  | 87.6  | 57.7  | 86.5# |
| 52  | 40.4  | 26.2  | 39.4# |




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM






#42  
1,2-Dichloroethane  
Concen: 19.28 ug/l  
RT: 8.531 min Scan# 2439  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 62 Resp: 141991

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 62  | 100   |       |       |
| 98  | 7.9   | 0.0   | 14.2  |







#43  
Isopropyl Acetate  
Concen: 19.62 ug/l  
RT: 8.560 min Scan# 2450  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 43 Resp: 202076  
Ion Ratio Lower Upper  
43 100  
61 25.2 19.0 28.4  
87 13.3 9.5 14.3




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#44  
Trichloroethene  
Concen: 19.51 ug/l  
RT: 9.217 min Scan# 2695  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion:130 Resp: 93674  
Ion Ratio Lower Upper  
130 100  
95 102.2 0.0 234.4









#45  
1,2-Dichloropropane  
Concen: 19.88 ug/l  
RT: 9.491 min Scan# 2797  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 63 Resp: 95565  
Ion Ratio Lower Upper  
63 100  
65 30.8 24.3 36.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#46  
Dibromomethane  
Concen: 18.96 ug/l  
RT: 9.577 min Scan# 2829  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26


Tgt Ion: 93 Resp: 70852  
Ion Ratio Lower Upper  
93 100  
95 84.3 66.1 99.1  
174 85.9 61.0 91.4





#47  
Bromodichloromethane  
Concen: 19.82 ug/l  
RT: 9.764 min Scan# 2899  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Tgt Ion: 83 Resp: 148773  
Ion Ratio Lower Upper  
83 100  
85 63.6 49.9 74.9  
127 8.7 6.1 9.1

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#48  
Methyl methacrylate  
Concen: 19.13 ug/l  
RT: 9.561 min Scan# 2823  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 41 Resp: 84162  
Ion Ratio Lower Upper  
41 100  
69 94.3 64.6 96.8  
39 86.5 53.3 79.9#









#49  
1,4-Dioxane  
Concen: 388.62 ug/l  
RT: 9.569 min Scan# 2826  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 88 Resp: 37210  
Ion Ratio Lower Upper  
88 100  
43 29.9 28.8 43.2  
58 59.6 60.4 90.6#

Instrument :  
MSVOA\_N  
ClientSampled :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#50  
Toluene-d8  
Concen: 52.50 ug/l  
RT: 10.443 min Scan# 3152  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 98 Resp: 1009655  
Ion Ratio Lower Upper  
98 100  
100 62.7 51.7 77.5









#51  
4-Methyl-2-Pentanone  
Concen: 94.18 ug/l  
RT: 10.330 min Scan# 3110  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 43 Resp: 617584  
Ion Ratio Lower Upper  
43 100  
58 37.5 28.6 43.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#52  
Toluene  
Concen: 19.17 ug/l  
RT: 10.507 min Scan# 3176  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 92 Resp: 239106  
Ion Ratio Lower Upper  
92 100  
91 181.6 139.6 209.4









#53  
t-1,3-Dichloropropene  
Concen: 19.67 ug/l  
RT: 10.719 min Scan# 3255  
Delta R.T. -0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 75 Resp: 150952  
Ion Ratio Lower Upper  
75 100  
77 29.8 25.1 37.7

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#54  
cis-1,3-Dichloropropene  
Concen: 19.66 ug/l  
RT: 10.191 min Scan# 3058  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 75 Resp: 155203  
Ion Ratio Lower Upper  
75 100  
77 29.3 24.6 37.0  
39 57.7 50.7 76.1








#55  
1,1,2-Trichloroethane  
Concen: 19.90 ug/l  
RT: 10.899 min Scan# 3322  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 97 Resp: 103523

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 84.4  | 72.0  | 108.0 |
| 85  | 53.4  | 46.8  | 70.2  |
| 99  | 58.2  | 49.8  | 74.6  |





#56  
Ethyl methacrylate  
Concen: 19.02 ug/l  
RT: 10.762 min Scan# 3271  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 69 Resp: 138946


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 66.3  | 58.6  | 88.0  |
| 39  | 54.2  | 38.3  | 57.5  |



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01


Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM



#57  
1,3-Dichloropropane  
Concen: 19.69 ug/l  
RT: 11.046 min Scan# 3377  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01



Tgt Ion: 76 Resp: 164295  
Ion Ratio Lower Upper  
76 100  
78 33.1 25.7 38.5




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#58  
2-Chloroethyl Vinyl ether  
Concen: 101.70 ug/l  
RT: 10.043 min Scan# 3003  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 63 Resp: 156981  
Ion Ratio Lower Upper  
63 100  
106 28.9 19.8 29.8









#59  
2-Hexanone  
Concen: 91.74 ug/l  
RT: 11.087 min Scan# 3392  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 43 Resp: 433872  
Ion Ratio Lower Upper  
43 100  
58 51.6 24.7 74.1



#60  
Dibromochloromethane  
Concen: 19.23 ug/l  
RT: 11.240 min Scan# 3449  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 129 Resp: 112161  
Ion Ratio Lower Upper  
129 100  
127 78.3 38.6 115.8



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01


Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM



#61  
1,2-Dibromoethane  
Concen: 18.86 ug/l  
RT: 11.347 min Scan# 3489  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Tgt Ion:107 Resp: 99778  
Ion Ratio Lower Upper  
107 100  
109 96.0 75.0 112.4

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#62  
4-Bromofluorobenzene  
Concen: 46.26 ug/l  
RT: 12.733 min Scan# 4006  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 95 Resp: 333436  
Ion Ratio Lower Upper  
95 100  
174 77.9 0.0 122.0  
176 74.6 0.0 117.6









#63  
Chlorobenzene-d5  
Concen: 50.00 ug/l  
RT: 11.746 min Scan# 3638  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:117 Resp: 709593  
Ion Ratio Lower Upper  
117 100  
82 54.2 53.1 79.7  
119 32.1 25.5 38.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#64  
Tetrachloroethene  
Concen: 19.54 ug/l  
RT: 10.979 min Scan# 3352  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:164 Resp: 82923  
Ion Ratio Lower Upper  
164 100  
166 128.5 102.1 153.1  
129 99.4 86.9 130.3  
131 97.8 83.6 125.4









#65  
Chlorobenzene  
Concen: 19.63 ug/l  
RT: 11.771 min Scan# 3647  
Delta R.T. -0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:112 Resp: 270482  
Ion Ratio Lower Upper  
112 100  
114 31.5 24.6 37.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.38 ug/l  
RT: 11.843 min Scan# 3674  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:131 Resp: 104638  
Ion Ratio Lower Upper  
131 100  
133 96.5 47.5 142.5  
119 67.4 33.1 99.3









#67  
Ethyl Benzene  
Concen: 19.26 ug/l  
RT: 11.846 min Scan# 3675  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 91 Resp: 475568  
Ion Ratio Lower Upper  
91 100  
106 31.4 23.4 35.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#68  
m/p-Xylenes  
Concen: 39.18 ug/l  
RT: 11.956 min Scan# 3716  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:106 Resp: 367806  
Ion Ratio Lower Upper  
106 100  
91 201.8 180.6 270.8









#69  
o-Xylene  
Concen: 19.60 ug/l  
RT: 12.283 min Scan# 3838  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:106 Resp: 183253  
Ion Ratio Lower Upper  
106 100  
91 212.5 118.3 354.9

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#70  
Styrene  
Concen: 19.05 ug/l  
RT: 12.296 min Scan# 3843  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:104 Resp: 296530  
Ion Ratio Lower Upper  
104 100  
78 50.6 48.2 72.2  
103 56.8 44.5 66.7









#71  
Bromoform  
Concen: 18.84 ug/l  
RT: 12.460 min Scan# 3904  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion:173 Resp: 74722  
Ion Ratio Lower Upper  
173 100  
175 45.9 23.9 71.8  
254 0.0 0.0 0.0



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.00 ug/l  
RT: 13.677 min Scan# 4358  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion:152 Resp: 304885  
Ion Ratio Lower Upper  
152 100  
115 62.4 36.1 108.5  
150 166.2 0.0 351.8



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM







#73  
Isopropylbenzene  
Concen: 20.15 ug/l  
RT: 12.581 min Scan# 3949  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 105 Resp: 485234  
Ion Ratio Lower Upper  
105 100  
120 25.6 12.5 37.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#74  
N-ethyl acetate  
Concen: 18.76 ug/l  
RT: 12.390 min Scan# 3878  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 43 Resp: 152926  
Ion Ratio Lower Upper  
43 100  
70 40.9 29.4 44.2  
55 25.3 17.0 25.4  
61 21.9 17.6 26.4









#75  
1,1,2,2-Tetrachloroethane  
Concen: 20.19 ug/l  
RT: 12.830 min Scan# 4042  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion: 83 Resp: 146532  
Ion Ratio Lower Upper  
83 100  
131 9.8 4.8 14.4  
85 65.2 32.2 96.6



#76  
1,2,3-Trichloropropane  
Concen: 19.83 ug/l m  
RT: 12.881 min Scan# 4061  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 75 Resp: 119188  
Ion Ratio Lower Upper  
75 100  
77 187.3 0.0 0.0#



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM







#77  
Bromobenzene  
Concen: 19.90 ug/l  
RT: 12.862 min Scan# 4054  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:156 Resp: 109127  
Ion Ratio Lower Upper  
156 100  
77 204.6 136.2 408.6  
158 96.9 48.4 145.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#78  
n-propylbenzene  
Concen: 20.45 ug/l  
RT: 12.924 min Scan# 4077  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 91 Resp: 560381  
Ion Ratio Lower Upper  
91 100  
120 22.4 10.4 31.4









#79  
2-Chlorotoluene  
Concen: 20.16 ug/l  
RT: 13.010 min Scan# 4109  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 91 Resp: 328935  
Ion Ratio Lower Upper  
91 100  
126 34.5 14.5 43.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 20.25 ug/l  
RT: 13.061 min Scan# 4128  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26


Tgt Ion:105 Resp: 403440  
Ion Ratio Lower Upper  
105 100  
120 47.8 22.2 66.6





#81  
trans-1,4-Dichloro-2-butene  
Concen: 19.32 ug/l  
RT: 12.629 min Scan# 3967  
Delta R.T. 0.003 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Tgt Ion: 75 Resp: 40073  
Ion Ratio Lower Upper  
75 100  
53 106.0 87.6 131.4  
89 45.1 37.8 56.8

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#82  
4-Chlorotoluene  
Concen: 20.11 ug/l  
RT: 13.106 min Scan# 4145  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion: 91 Resp: 325586  
Ion Ratio Lower Upper  
91 100  
126 33.2 14.8 44.4





#83  
tert-Butylbenzene  
Concen: 19.91 ug/l  
RT: 13.326 min Scan# 4227  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01



Tgt Ion:119 Resp: 350463  
Ion Ratio Lower Upper  
119 100  
91 68.7 39.5 118.5  
134 24.0 12.0 36.0




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#84  
1,2,4-Trimethylbenzene  
Concen: 20.61 ug/l  
RT: 13.369 min Scan# 4243  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26


Tgt Ion:105 Resp: 403028  
Ion Ratio Lower Upper  
105 100  
120 44.0 20.6 61.8





#85  
 sec-Butylbenzene  
 Concen: 20.28 ug/l  
 RT: 13.503 min Scan# 4293  
 Delta R.T. 0.000 min  
 Lab File: VN068821.D  
 Acq: 6 Oct 2021 16:26




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1006WBS01



Tgt Ion:105 Resp: 508024  
 Ion Ratio Lower Upper  
 105 100  
 134 19.7 9.0 27.1




Manual Integrations  
 APPROVED


MMDadoda  
 10/7/2021 5:34:27 PM



#86  
 p-Isopropyltoluene  
 Concen: 20.35 ug/l  
 RT: 13.618 min Scan# 4336  
 Delta R.T. -0.001 min  
 Lab File: VN068821.D  
 Acq: 6 Oct 2021 16:26

Tgt Ion:119 Resp: 412740  
 Ion Ratio Lower Upper  
 119 100  
 134 25.5 12.4 37.4  
 91 24.6 14.8 44.3









#87  
1,3-Dichlorobenzene  
Concen: 19.72 ug/l  
RT: 13.618 min Scan# 4336  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:146 Resp: 202981  
Ion Ratio Lower Upper  
146 100  
111 41.4 24.3 72.8  
148 64.0 32.0 96.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM



#88  
1,4-Dichlorobenzene  
Concen: 19.46 ug/l  
RT: 13.696 min Scan# 4365  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26


Tgt Ion:146 Resp: 197655  
Ion Ratio Lower Upper  
146 100  
111 41.9 23.4 70.2  
148 64.7 31.8 95.3





#89  
n-Butylbenzene  
Concen: 19.56 ug/l  
RT: 13.943 min Scan# 4457  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Tgt Ion: 91 Resp: 334738  
Ion Ratio Lower Upper  
91 100  
92 49.4 25.7 77.1  
134 24.8 11.2 33.6

Manual Integrations  
APPROVED



MMDadoda  
10/7/2021 5:34:27 PM




#90  
Hexachloroethane  
Concen: 20.38 ug/l  
RT: 14.211 min Scan# 4557  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion: 117 Resp: 79901  
Ion Ratio Lower Upper  
117 100  
201 66.5 33.5 100.5









#91  
1,2-Dichlorobenzene  
Concen: 20.00 ug/l  
RT: 13.991 min Scan# 4475  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion:146 Resp: 198136  
Ion Ratio Lower Upper  
146 100  
111 42.4 24.9 74.6  
148 63.0 31.6 95.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Manual Integrations  
APPROVED


MMDadoda  
10/7/2021 5:34:27 PM




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 20.18 ug/l  
RT: 14.603 min Scan# 4703  
Delta R.T. -0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26

Tgt Ion: 75 Resp: 25016  
Ion Ratio Lower Upper  
75 100  
155 77.2 29.8 89.5  
157 98.9 37.6 112.8









#93  
1,2,4-Trichlorobenzene  
Concen: 18.10 ug/l  
RT: 15.265 min Scan# 4950  
Delta R.T. 0.000 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26





Tgt Ion:180 Resp: 82083  
Ion Ratio Lower Upper  
180 100  
182 96.5 47.5 142.5  
145 32.6 17.2 51.6



#94  
Hexachlorobutadiene  
Concen: 20.46 ug/l  
RT: 15.372 min Scan# 4990  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Tgt Ion:225 Resp: 49568  
Ion Ratio Lower Upper  
225 100  
223 64.5 31.4 94.0  
227 63.2 31.2 93.6



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01


Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM



#95  
Naphthalene  
Concen: 16.59 ug/l  
RT: 15.509 min Scan# 5041  
Delta R.T. 0.002 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1006WBS01




Tgt Ion:128 Resp: 200732  
Ion Ratio Lower Upper  
128 100  
127 13.8 10.6 15.8  
129 10.7 8.6 12.8

Manual Integrations  
APPROVED

MMDadoda  
10/7/2021 5:34:27 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 18.80 ug/l  
RT: 15.702 min Scan# 5113  
Delta R.T. -0.001 min  
Lab File: VN068821.D  
Acq: 6 Oct 2021 16:26



Tgt Ion:180 Resp: 78390  
Ion Ratio Lower Upper  
180 100  
182 96.1 47.8 143.3  
145 34.8 18.2 54.6

