

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074777.D
 Acq On : 06 Oct 2022 12:39
 Operator : JC\MD
 Sample : N5005-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-100422

Quant Time: Oct 06 12:57:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	241293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	452560	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	468800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	221828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	318488	55.707	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.420%
35) Dibromofluoromethane	8.183	113	245843	53.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.180%
50) Toluene-d8	10.571	98	1009582	51.457	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.920%
62) 4-Bromofluorobenzene	12.853	95	256063	42.934	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074777.D
Acq On : 06 Oct 2022 12:39
Operator : JC\MD
Sample : N5005-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-100422

Quant Time: Oct 06 12:57:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

