

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074779.D
 Acq On : 06 Oct 2022 13:27
 Operator : JC\MD
 Sample : N4983-03DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-763-765DL

Quant Time: Oct 06 16:30:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	252354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	469876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	465916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	221886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	320348	53.576	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.160%	
35) Dibromofluoromethane	8.177	113	259007	54.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.120%	
50) Toluene-d8	10.571	98	1027732	50.452	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.900%	
62) 4-Bromofluorobenzene	12.853	95	271246	43.803	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.600%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.389	101	13138	2.855	ug/l	# 71
44) Trichloroethene	9.359	130	276925	53.450	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074779.D
Acq On : 06 Oct 2022 13:27
Operator : JC\MD
Sample : N4983-03DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-763-765DL

Quant Time: Oct 06 16:30:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

