

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074783.D
 Acq On : 06 Oct 2022 15:03
 Operator : JC\MD
 Sample : N4910-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/07/2022
 Supervised By : Mahesh Dadoda 10/07/2022

Quant Time: Oct 06 19:04:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	234276	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	439001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	445329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	216874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	320006	57.649	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.300%			
35) Dibromofluoromethane	8.177	113	242499	54.673	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.340%			
50) Toluene-d8	10.571	98	989963	52.015	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.040%			
62) 4-Bromofluorobenzene	12.853	95	256877	44.400	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 88.800%			
Target Compounds						
						Qvalue
44) Trichloroethene	9.353	130	2846m	0.588	ug/l	
64) Tetrachloroethene	11.106	164	35018	9.686	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074783.D
Acq On : 06 Oct 2022 15:03
Operator : JC\MD
Sample : N4910-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/07/2022
Supervised By : Mahesh Dadoda 10/07/2022

Quant Time: Oct 06 19:04:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

