

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074784.D
 Acq On : 06 Oct 2022 15:27
 Operator : JC\MD
 Sample : N4910-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Quant Time: Oct 06 19:04:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	242666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	456120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	461458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215181	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	314175	54.641	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.280%	
35) Dibromofluoromethane	8.177	113	245460	53.263	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.520%	
50) Toluene-d8	10.570	98	1012302	51.193	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.380%	
62) 4-Bromofluorobenzene	12.853	95	254630	42.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.720%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	14668	2.917	ug/l	93
64) Tetrachloroethene	11.106	164	137390	36.676	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074784.D
Acq On : 06 Oct 2022 15:27
Operator : JC\MD
Sample : N4910-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Quant Time: Oct 06 19:04:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

