

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074787.D
 Acq On : 06 Oct 2022 16:40
 Operator : JC\MD
 Sample : N4910-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Oct 06 19:16:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	245707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	465735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	473256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	223333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	312350	53.652	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.300%	
35) Dibromofluoromethane	8.183	113	257652	54.755	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.500%	
50) Toluene-d8	10.571	98	1023270	50.679	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	12.853	95	272906	44.463	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.920%	
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	7.494	96	14925	2.458	ug/l	92
44) Trichloroethene	9.359	130	6679	1.301	ug/l	90
64) Tetrachloroethene	11.106	164	688627	179.243	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 06 19:16:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

