

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074789.D
 Acq On : 06 Oct 2022 17:28
 Operator : JC\MD
 Sample : N4910-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Oct 06 19:17:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	232281	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.112	114	452140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	455437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219066	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	318694	57.905	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.820%
35) Dibromofluoromethane	8.177	113	248189	54.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.660%
50) Toluene-d8	10.570	98	992563	50.636	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.853	95	263404	44.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074789.D
Acq On : 06 Oct 2022 17:28
Operator : JC\MD
Sample : N4910-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Oct 06 19:17:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

