

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074790.D
 Acq On : 06 Oct 2022 17:52
 Operator : JC\MD
 Sample : N4910-06RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5RE

Quant Time: Oct 06 19:18:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	281904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	542980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	535962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	244462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	335789	50.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.540%
35) Dibromofluoromethane	8.177	113	264551	48.223	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.440%
50) Toluene-d8	10.571	98	1071444	45.516	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.040%
62) 4-Bromofluorobenzene	12.853	95	266521	37.246	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	74.500%#
Target Compounds						
						Qvalue
44) Trichloroethene	9.359	130	11069	1.849	ug/l	66
64) Tetrachloroethene	11.106	164	90956	20.905	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074790.D
Acq On : 06 Oct 2022 17:52
Operator : JC\MD
Sample : N4910-06RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5RE

Quant Time: Oct 06 19:18:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

