

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074793.D
 Acq On : 06 Oct 2022 19:05
 Operator : JC\MD
 Sample : N4910-02DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1DL

Quant Time: Oct 06 19:33:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.241	168	228376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	422448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	426781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	207992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	316327	58.458	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.920%
35) Dibromofluoromethane	8.177	113	252994	59.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.540%
50) Toluene-d8	10.571	98	985549	53.812	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.620%
62) 4-Bromofluorobenzene	12.853	95	259391	46.592	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.180%
Target Compounds						
						Qvalue
44) Trichloroethene	9.359	130	1541	0.331	ug/l	70
64) Tetrachloroethene	11.106	164	148117	42.752	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074793.D
Acq On : 06 Oct 2022 19:05
Operator : JC\MD
Sample : N4910-02DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1DL

Quant Time: Oct 06 19:33:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

