

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074794.D
 Acq On : 06 Oct 2022 19:29
 Operator : JC\MD
 Sample : N4910-03RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2RE

Quant Time: Oct 06 19:49:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	244769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	455345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	448339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	216809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	309177	53.310	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.620%	
35) Dibromofluoromethane	8.177	113	240275	52.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.460%	
50) Toluene-d8	10.571	98	981130	49.701	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	12.853	95	253425	42.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074794.D
Acq On : 06 Oct 2022 19:29
Operator : JC\MD
Sample : N4910-03RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2RE

Quant Time: Oct 06 19:49:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

