

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074796.D
 Acq On : 06 Oct 2022 20:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 06:59:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	1.025	1.012	1.3	82	0.00
3 P	Chloromethane	1.525	1.358	11.0	81	0.00
4 C	Vinyl Chloride	1.888	1.963	-4.0#	88	0.00
5 T	Bromomethane	1.331	1.411	-6.0	87	0.00
6 T	Chloroethane	1.397	1.445	-3.4	86	0.00
7 T	Trichlorofluoromethane	1.488	1.378	7.4	84	0.00
8 T	Diethyl Ether	0.776	0.668	13.9	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.912	0.840	7.9	81	0.00
10 T	Methyl Iodide	1.102	1.043	5.4	80	0.00
11 T	Tert butyl alcohol	0.397	0.345	13.1	78	-0.01
12 CM	1,1-Dichloroethene	0.970	0.837	13.7#	79	0.00
13 T	Acrolein	0.112	0.100	10.7	82	0.00
14 T	Allyl chloride	1.994	1.746	12.4	73	0.00
15 T	Acrylonitrile	0.865	0.815	5.8	79	0.00
16 T	Acetone	0.671	0.636	5.2	82	0.00
17 T	Carbon Disulfide	2.652	2.296	13.4	76	0.00
18 T	Methyl Acetate	2.486	2.394	3.7	84	0.00
19 T	Methyl tert-butyl Ether	3.775	3.632	3.8	81	0.00
20 T	Methylene Chloride	1.263	1.084	14.2	80	0.00
21 T	trans-1,2-Dichloroethene	1.038	0.940	9.4	80	0.00
22 T	Diisopropyl ether	4.236	4.192	1.0	84	0.00
23 T	Vinyl Acetate	3.479	3.607	-3.7	82	0.00
24 P	1,1-Dichloroethane	2.184	2.066	5.4	80	0.00
25 T	2-Butanone	1.218	1.179	3.2	82	0.00
26 T	2,2-Dichloropropane	1.801	1.523	15.4	74	0.00
27 T	cis-1,2-Dichloroethene	1.236	1.185	4.1	82	0.01
28 T	Bromochloromethane	0.985	1.026	-4.2	82	0.00
29 T	Tetrahydrofuran	0.849	0.816	3.9	81	0.00
30 C	Chloroform	2.000	2.008	-0.4#	84	0.00
31 T	Cyclohexane	2.048	1.875	8.4	79	0.00
32 T	1,1,1-Trichloroethane	1.705	1.680	1.5	82	0.00
33 S	1,2-Dichloroethane-d4	1.185	1.266	-6.8	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.505	0.537	-6.3	88	0.00
36 T	1,1-Dichloropropene	0.818	0.835	-2.1	83	0.00
37 T	Ethyl Acetate	3.280	3.292	-0.4	82	0.00
38 T	Carbon Tetrachloride	0.771	0.787	-2.1	83	0.00
39 T	Methylcyclohexane	1.059	1.026	3.1	76	0.00
40 TM	Benzene	2.647	2.689	-1.6	82	0.00
41 T	Methacrylonitrile	0.646	0.661	-2.3	83	0.00
42 TM	1,2-Dichloroethane	0.831	0.850	-2.3	81	0.00
43 T	Isopropyl Acetate	2.027	2.094	-3.3	84	0.00
44 TM	Trichloroethene	0.551	0.556	-0.9	82	0.00
45 C	1,2-Dichloropropane	0.717	0.713	0.6#	79	0.00
46 T	Dibromomethane	0.421	0.444	-5.5	81	0.00
47 T	Bromodichloromethane	0.859	0.892	-3.8	83	0.00
48 T	Methyl methacrylate	0.832	0.888	-6.7	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
 Data File : VN074796.D
 Acq On : 06 Oct 2022 20:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 06:59:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.018	0.017	5.6	78	0.00
50 S	Toluene-d8	2.168	2.333	-7.6	86	0.00
51 T	4-Methyl-2-Pentanone	1.305	1.379	-5.7	83	0.00
52 CM	Toluene	1.590	1.712	-7.7#	83	0.00
53 T	t-1,3-Dichloropropene	1.024	1.079	-5.4	81	0.00
54 T	cis-1,3-Dichloropropene	1.105	1.101	0.4	79	0.00
55 T	1,1,2-Trichloroethane	0.639	0.673	-5.3	82	0.00
56 T	Ethyl methacrylate	1.165	1.220	-4.7	81	0.00
57 T	1,3-Dichloropropane	1.152	1.189	-3.2	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.420	0.423	-0.7	73	0.00
59 T	2-Hexanone	0.987	1.053	-6.7	83	0.00
60 T	Dibromochloromethane	0.630	0.681	-8.1	83	0.00
61 T	1,2-Dibromoethane	0.650	0.681	-4.8	83	0.00
62 S	4-Bromofluorobenzene	0.659	0.744	-12.9	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.406	0.400	1.5	84	0.00
65 PM	Chlorobenzene	1.532	1.541	-0.6	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.559	0.563	-0.7	84	0.00
67 C	Ethyl Benzene	2.855	2.970	-4.0#	82	0.00
68 T	m/p-Xylenes	1.072	1.139	-6.2	83	0.00
69 T	o-Xylene	1.100	1.150	-4.5	82	0.00
70 T	Styrene	1.733	1.890	-9.1	83	0.00
71 P	Bromoform	0.427	0.452	-5.9	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	4.811	4.502	6.4	83	0.00
74 T	N-amyl acetate	2.909	2.580	11.3	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.915	1.664	13.1	85	0.00
76 T	1,2,3-Trichloropropane	1.492	1.274	14.6	73	0.00
77 T	Bromobenzene	1.013	0.926	8.6	85	0.00
78 T	n-propylbenzene	5.543	5.299	4.4	81	0.00
79 T	2-Chlorotoluene	3.361	3.124	7.1	84	0.00
80 T	1,3,5-Trimethylbenzene	3.837	3.724	2.9	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.766	0.610	20.4	70	0.00
82 T	4-Chlorotoluene	3.115	2.999	3.7	81	0.00
83 T	tert-Butylbenzene	3.316	3.222	2.8	82	0.00
84 T	1,2,4-Trimethylbenzene	3.880	3.810	1.8	84	0.00
85 T	sec-Butylbenzene	4.819	4.821	-0.0	83	0.00
86 T	p-Isopropyltoluene	3.719	3.798	-2.1	83	0.00
87 T	1,3-Dichlorobenzene	1.908	1.819	4.7	83	0.00
88 T	1,4-Dichlorobenzene	1.802	1.795	0.4	84	0.00
89 T	n-Butylbenzene	3.278	3.376	-3.0	84	0.00
90 T	Hexachloroethane	0.831	0.736	11.4	79	0.00
91 T	1,2-Dichlorobenzene	1.838	1.822	0.9	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.399	0.344	13.8	83	0.00
93 T	1,2,4-Trichlorobenzene	0.804	0.822	-2.2	86	0.00
94 T	Hexachlorobutadiene	0.376	0.356	5.3	90	0.00
95 T	Naphthalene	3.475	3.756	-8.1	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100622\
Data File : VN074796.D
Acq On : 06 Oct 2022 20:17
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 07 06:59:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.875	0.825	5.7	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6