

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079436.D
 Acq On : 06 Oct 2023 10:21
 Operator : JC\MD
 Sample : VN1006WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1006WBL01

Quant Time: Oct 06 23:00:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	219838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	408553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	351961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	94388	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	204403	57.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.680%
35) Dibromofluoromethane	8.171	113	157364	53.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.200%
50) Toluene-d8	10.571	98	546261	51.421	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.853	95	136071	39.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.820%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
Data File : VN079436.D
Acq On : 06 Oct 2023 10:21
Operator : JC\MD
Sample : VN1006WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1006WBL01

Quant Time: Oct 06 23:00:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

