

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079445.D
 Acq On : 06 Oct 2023 14:18
 Operator : JC\MD
 Sample : 04689-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TBRE

Quant Time: Oct 06 23:03:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	198267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	328731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	84107	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	184010	57.239	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.480%
35) Dibromofluoromethane	8.171	113	139629	49.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	10.571	98	477737	46.414	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.820%
62) 4-Bromofluorobenzene	12.853	95	117747	35.646	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	71.300%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
Data File : VN079445.D
Acq On : 06 Oct 2023 14:18
Operator : JC\MD
Sample : 04689-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TBRE

Quant Time: Oct 06 23:03:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

