

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079448.D
 Acq On : 06 Oct 2023 15:30
 Operator : JC\MD
 Sample : PB156147ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156147ZHE#03

Quant Time: Oct 06 23:04:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	152063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	322398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	314619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	77620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136697	55.442	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.880%		
35) Dibromofluoromethane	8.171	113	116430	50.253	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.500%		
50) Toluene-d8	10.571	98	422263	50.371	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.740%		
62) 4-Bromofluorobenzene	12.853	95	112554	41.837	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 83.680%		
Target Compounds						
						Qvalue
16) Acetone	4.447	43	33410	34.164	ug/l #	86
20) Methylene Chloride	5.289	84	44689	20.513	ug/l #	83
43) Isopropyl Acetate	8.694	43	234820	38.737	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
Data File : VN079448.D
Acq On : 06 Oct 2023 15:30
Operator : JC\MD
Sample : PB156147ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156147ZHE#03

Quant Time: Oct 06 23:04:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

