

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079449.D
 Acq On : 06 Oct 2023 15:54
 Operator : JC\MD
 Sample : PB156147ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156147ZHE#04

Quant Time: Oct 06 23:05:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156106	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	337678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	345849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	94672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	145423	57.454	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.900%
35) Dibromofluoromethane	8.171	113	119189	49.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	10.571	98	446739	50.879	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.760%
62) 4-Bromofluorobenzene	12.853	95	126295	44.820	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.640%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	33592	33.460	ug/l #	88
20) Methylene Chloride	5.283	84	44777	20.006	ug/l #	84
37) Ethyl Acetate	7.577	43	4368	1.205	ug/l #	78
43) Isopropyl Acetate	8.694	43	241609	38.054	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
Data File : VN079449.D
Acq On : 06 Oct 2023 15:54
Operator : JC\MD
Sample : PB156147ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156147ZHE#04

Quant Time: Oct 06 23:05:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

