

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079454.D
 Acq On : 06 Oct 2023 17:56
 Operator : JC\MD
 Sample : 04731-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DTP-1

Quant Time: Oct 06 23:06:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	186340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	404827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	361590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	94536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	206626	68.388	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 136.780%#	
35) Dibromofluoromethane	8.171	113	148820	51.154	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.300%	
50) Toluene-d8	10.570	98	541082	51.402	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.800%	
62) 4-Bromofluorobenzene	12.853	95	136946	40.539	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 81.080%#	
Target Compounds						
						Qvalue
16) Acetone	4.441	43	23518	19.625	ug/l #	87
20) Methylene Chloride	5.277	84	15152	5.235	ug/l #	84
43) Isopropyl Acetate	8.694	43	327626	43.042	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
Data File : VN079454.D
Acq On : 06 Oct 2023 17:56
Operator : JC\MD
Sample : 04731-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DTP-1

Quant Time: Oct 06 23:06:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

