

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079455.D
 Acq On : 06 Oct 2023 18:20
 Operator : JC\MD
 Sample : 04731-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DTP-2

Quant Time: Oct 06 23:07:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	186275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	396958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	354573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	88257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	196450	65.043	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	130.080%#
35) Dibromofluoromethane	8.171	113	144573	50.680	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.360%
50) Toluene-d8	10.570	98	518235	50.208	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.420%
62) 4-Bromofluorobenzene	12.853	95	127197	38.399	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	76.800%#
Target Compounds						
						Qvalue
16) Acetone	4.441	43	23980	20.017	ug/l #	87
18) Methyl Acetate	5.041	43	7038	1.200	ug/l #	70
20) Methylene Chloride	5.288	84	13381	4.554	ug/l	91
37) Ethyl Acetate	7.577	43	5715	1.341	ug/l #	79
43) Isopropyl Acetate	8.694	43	312630	41.887	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100623\
 Data File : VN079455.D
 Acq On : 06 Oct 2023 18:20
 Operator : JC\MD
 Sample : 04731-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DTP-2

Quant Time: Oct 06 23:07:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

