

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN100719\
 Data File : VN058573.D
 Acq On : 7 Oct 2019 17:53
 Operator : JC/SP
 Sample : K5197-03DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-28B_100219DL

Quant Time: Oct 17 04:30:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	505623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	793829	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	695958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	285079	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	340816	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
35) Dibromofluoromethane	7.57	113	239110	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	10.08	98	967602	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.41	95	331165	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
Target Compounds						
44) Trichloroethene	8.83	130	5674	1.053	ug/l	99
64) Tetrachloroethene	10.63	164	354757	80.232	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN100719\
Data File : VN058573.D
Acq On : 7 Oct 2019 17:53
Operator : JC/SP
Sample : K5197-03DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-28B_100219DL

Quant Time: Oct 17 04:30:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

