

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100719\
 Data File : VN058577.D
 Acq On : 7 Oct 2019 19:21
 Operator : JC/SP
 Sample : PB123664TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123664TB

Quant Time: Oct 08 15:27:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	496072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	789527	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	703419	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	293364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	338707	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
35) Dibromofluoromethane	7.57	113	241889	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	10.08	98	986411	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	12.41	95	337988	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
Target Compounds						
16) Acetone	3.80	43	16038	6.490	ug/l	94
20) Methylene Chloride	4.53	84	41876	7.343	ug/l	93
43) Isopropyl Acetate	8.15	43	107910	8.965	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100719\
Data File : VN058577.D
Acq On : 7 Oct 2019 19:21
Operator : JC/SP
Sample : PB123664TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123664TB

Quant Time: Oct 08 15:27:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

