

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100719\
 Data File : VN058582.D
 Acq On : 7 Oct 2019 21:12
 Operator : JC/SP
 Sample : K5230-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Oct 08 15:40:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	480959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	763703	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	669970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	279727	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	328464	46.73	ug/l	0.00
Spiked Amount			Recovery	=	93.46%	
35) Dibromofluoromethane	7.57	113	236863	50.60	ug/l	0.00
Spiked Amount			Recovery	=	101.20%	
50) Toluene-d8	10.08	98	949379	51.78	ug/l	0.00
Spiked Amount			Recovery	=	103.56%	
62) 4-Bromofluorobenzene	12.41	95	317675	47.91	ug/l	0.00
Spiked Amount			Recovery	=	95.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.80	43	32050	13.377	ug/l	99
20) Methylene Chloride	4.53	84	96435	17.441	ug/l	95
25) 2-Butanone	6.83	43	4522	1.375	ug/l	94
43) Isopropyl Acetate	8.15	43	83922	7.208	ug/l	99
95) Naphthalene	15.13	128	175487	12.819	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100719\
Data File : VN058582.D
Acq On : 7 Oct 2019 21:12
Operator : JC/SP
Sample : K5230-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Oct 08 15:40:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

