

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063935.D
 Acq On : 7 Oct 2020 11:20
 Operator : JC/MD
 Sample : VN1007WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1007WBL01

Quant Time: Oct 08 05:16:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	227470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	410268	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	427178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	172241	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200385	53.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.64%	
35) Dibromofluoromethane	7.55	113	144743	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
50) Toluene-d8	10.06	98	535345	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
62) 4-Bromofluorobenzene	12.38	95	202641	49.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063935.D
Acq On : 7 Oct 2020 11:20
Operator : JC/MD
Sample : VN1007WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1007WBL01

Quant Time: Oct 08 05:16:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

