

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063938.D
 Acq On : 7 Oct 2020 12:56
 Operator : JC/MD
 Sample : L4253-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 08 05:22:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	228206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	416748	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	435412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169614	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	202401	53.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.38%	
35) Dibromofluoromethane	7.55	113	145688	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
50) Toluene-d8	10.06	98	544747	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
62) 4-Bromofluorobenzene	12.38	95	205897	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063938.D
Acq On : 7 Oct 2020 12:56
Operator : JC/MD
Sample : L4253-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 08 05:22:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

