

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063945.D
 Acq On : 7 Oct 2020 16:07
 Operator : JC/MD
 Sample : L4293-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-7-MW-3

Manual Integrations
 APPROVED

MMDadoda
 10/8/2020 10:08:48 AM

Quant Time: Oct 08 05:38:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	228817	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414463	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	433634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	204046	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
35) Dibromofluoromethane	7.55	113	144125	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	10.06	98	532631	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.38	95	205698	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
Target Compounds						
16) Acetone	3.78	43	8766	7.533	ug/l	Qvalue 100
20) Methylene Chloride	4.52	84	1945m	0.826	ug/l	
95) Naphthalene	15.11	128	16779	3.596	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063945.D
Acq On : 7 Oct 2020 16:07
Operator : JC/MD
Sample : L4293-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-7-MW-3

Manual Integrations
APPROVED

MMDadoda
10/8/2020 10:08:48 AM

Quant Time: Oct 08 05:38:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

