

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063946.D
 Acq On : 7 Oct 2020 16:31
 Operator : JC/MD
 Sample : L4293-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-17-MW-4

Quant Time: Oct 08 05:40:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	233079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	424507	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	175977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	205327	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
35) Dibromofluoromethane	7.55	113	149135	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
50) Toluene-d8	10.06	98	543838	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	12.38	95	211014	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
Target Compounds						
16) Acetone	3.78	43	4371	3.688	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063946.D
Acq On : 7 Oct 2020 16:31
Operator : JC/MD
Sample : L4293-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-17-MW-4

Quant Time: Oct 08 05:40:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

