

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063953.D
 Acq On : 7 Oct 2020 19:19
 Operator : JC/MD
 Sample : L4253-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 08 06:09:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	251057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	450340	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	463622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191886	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	212411	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
35) Dibromofluoromethane	7.55	113	152728	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
50) Toluene-d8	10.06	98	571118	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
62) 4-Bromofluorobenzene	12.38	95	227959	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063953.D
Acq On : 7 Oct 2020 19:19
Operator : JC/MD
Sample : L4253-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 08 06:09:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

