

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063954.D
 Acq On : 7 Oct 2020 19:43
 Operator : JC/MD
 Sample : L4260-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 08 06:12:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	243823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	442475	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	454414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	208551	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
35) Dibromofluoromethane	7.55	113	150820	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	10.06	98	561194	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	12.38	95	222509	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
Target Compounds						
68) m/p-Xylenes	11.60	106	2775	0.439	ug/l #	48
84) 1,2,4-Trimethylbenzene	13.02	105	4575	0.392	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063954.D
Acq On : 7 Oct 2020 19:43
Operator : JC/MD
Sample : L4260-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 08 06:12:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

